



नेपाल प्रहरी अस्पताल

(प्रबन्ध शाखा)

पत्र संख्या: २०७७/०७८

प्रा.प्र.सं.

च.नं. १६८२

फोन: { ०१-४४१२४३०
०१-४४१२५३०

Ext.: { ०१-४४१२५३०
१२९

पो.व.नं. २११४७

फ्याक्स: ०१-४३५१५३

महाराजगंज, काठमाण्डौ ।

मिति: २०७७/०७/२४ गते ।

विषय : बोलपत्र सम्बन्धी कागजातहरूको संशोधन सम्बन्धमा ।

विश्व व्यापी रूपमा फैलिएको महामारी Coronavirus (COVID -19) को प्रकोप दैनिक रूपमा बढ्दै गईरहेकोले प्रत्यक्ष रूपमा खटिने नेपाल प्रहरीका सुरक्षाकर्मीहरू लगायत अन्य बिरामीहरू समेतको स्वास्थ्य हितलाई ध्यानमा राखि धनजनको क्षतिलाई ब्यबस्थापन एवं न्युनिकरण गर्न तालुक स्वास्थ्य तथा जनसंख्या मन्त्रालय, नीति तथा योजना शाखाको च.नं. २५८ मिति २०७७/०७/०५ गतेको पत्र अनुसार कोभिड-१९ महामारीका Moderate Service Cases हरूको ICU र General Bed को Service बीचको अवस्थाको सेवा उपलब्ध गराउन आवश्यक सुबिधा र उपकरण सहितका High Dependency Unit (HDU) १५ दिनभित्र स्तरोन्नती गर्नको लागि नेपाल सरकार (मन्त्रिपरिषद) को मिति २०७७/०६/२६ मा निर्णय भए बमोजिम High Dependency Unit (HDU) को २ युनिट (५० बेड) यस अस्पतालमा तत्काल तयार गर्नु पर्ने भएकोले सार्वजनिक खरिद ऐन २०६३ (संसोधन सहित) को दफा ६६ र सार्वजनिक खरिद नियमावली २०६४ (संसोधन सहित) को नियम १४५ मा ब्यवस्था भए बमोजिम उक्त वार्डको स्थापनाको लागि आवश्यक मेशिनरी उपकरण लगायतका सामाग्रीहरू तथा यस अस्पतालमा कोभिड-१९ माहामारीबाट संक्रमित बिरामीहरूको उपचारको लागि अति आवश्यक बिभिन्न मेडिकल मेशिनरी तथा औजारहरू लगायतका सामाग्रीहरू समेत विशेष परिस्थितिमा आकस्मिक रूपमा तत्काल खरिद गर्नको लागि मिति २०७७/०७/२१ गतेको गोरखापत्र राष्ट्रिय दैनिकमा प्रकाशित सूचना बमोजिमको बोलपत्र सम्बन्धी कागजातहरूमा मिति २०७७/०७/२४ गते १३:०० बजे यस अस्पतालमा आयोजना गरिएको बोलपत्र पेश गर्नु पूर्वको बैठक (Pre-Bid Meeting) मा भएको छलफल तथा बोलपत्रदाताहरूको तर्फबाट प्राप्त भएको सुझाव समेतलाई ध्यान दिई देहाय वमोजिम संशोधन/परिमार्जन गरिएकोले सम्बन्धित सवैको जानकारीको लागि अनुरोध गरिएको छ ।

साथै बोलपत्र दाखिला गर्ने तथा बोलपत्र खोल्ने मिति र समय यथावत रहेको व्यहोरा समेत जानकारीको लागि अनुरोध गरिएको छ ।

निम्न:

संशोधन गरिएको मेशिनरी उपकरणको सि.नं.	मेशिनरी उपकरणको नाम	बिबरण		संशोधन तथा थपको भएको बिबरण
		बुँदा नं.	अस्पतालले निर्धारण गरेको स्पेशिफिकेशन	
1	ICU Bed and Accessories	4.1	Movement Shall be : Backrest lifting : 0° to 65° or better Knee break lifting : 0° to 30° or better Trendlenburg and reverse trendlenburg : 0° to 10° or more Motorized height adjustment range : 500mm(min)-800mm(max)	Movement Shall be : Backrest lifting : 0° to 65° or better Knee break lifting : 0° to 30° or better Trendlenburg and reverse trendlenburg : 0° to 10° or more Motorized height adjustment range : 350 mm (min)-800 mm(max)



hsh

पाना नं. २

संशोधन गरिएको मेशिनरी उपकरणको सि.नं.	मेशिनरी उपकरणको नाम	बिबरण		
		बुँदा नं.	अस्पतालले निर्धारण गरेको स्पेशिफिकेशन	
		9	Must submit valid CE (93/42 EEC Directives) approved product certificate or US-FDA approved product certificate must be valid.	संशोधन तथा थपको भएको बिबरण
			Must submit valid CE (93/42 EEC Directives) approved product certificate or US-FDA approved or Japanese FDA product certificate must be valid.	
5	Patient Monitor	3.1	Patient monitor with ECG, Resp., SpO2, NIBP, 2-ch x Temp, 2-ch x IBP and EtCO2. It should have Cardiac output, anesthetic gas module and BIS in optional. It shall be clearly mention in the brochure.	Patient monitor with ECG, Resp., SpO2, NIBP, 2-ch x Temp, 2-ch x IBP and EtCO2. It should have Cardiac output in optional. It shall be clearly mention in the brochure.
		4.6	With powerful data storage for up to 48 hours of full disclosure, 120 hours or more of tabular and graphic trends and 100 or more NIBP measurements.	Not required
		4.8	Should have up to 1000 or more alarm events with associated waveforms	Not required
13	Portable Colour Doppler Ultrasound Machine	3.1	System shall come with main unit, 2 probes, 1 unit of black and white video thermal printer and Mobile Trolley.	System shall come with main unit, 3 probes, 1 unit of black and white video thermal printer and Mobile Trolley.
		4.1	Main applications: OB/GYN, abdominal, cardiac and small parts.	Main applications: OB/GYN, abdominal, cardiac and small parts. All the application's software need to be provided.
14	Video Laryngoscope	4.1	For disposable blade should supply 50 nos. of blades compatible for both adult and paediatric along with each unit.	Reusable blades 3 sets for infant, adult and paediatric.
		7.2	CE or USFDA approved product certificate.	CE and USFDA approved product certificate.

hsh




पाना नं. 3

संशोधन गरिएको मेशिनरी उपकरणको सि.नं.	मेशिनरी उपकरणको नाम	अस्पतालले निर्धारण गरेको स्पेशिफिकेशन		संशोधन तथा थपको भएको बिबरण
		बुँदा नं.		
15	Non-Invasive Cardiac Output Monitor	9	a. Transducer as mentioned	a. Transducer as mentioned 2 sets
		11.1		The bidder shall provide high quality principal company made comfortable movable trolley.
		31	-FDA or CE (Certification no. should be provided) approved to meet the international Standard and requirements for safety. Valid certificate should be attached	USFDA and CE (Certification no. should be provided) approved to meet the international Standard and requirements for safety. Valid certificate should be attached
16	Portable X-ray Machine	5.1		Lead apron two sets.

कार्यार्थ एवं बोधार्थ:

श्री प्रि-विड मिटिङमा सहभागी सम्पुर्ण आपूर्तिकर्ताहरु ।
श्री नेपाल प्रहरी अस्पताल, कम्प्यूटर शाखा: यस अस्पतालको
Website मा राखिदिन हुन ।
श्री नेपाल प्रहरी अस्पताल, सूचना पाटी ।


मेडिकल डाइरेक्टर
(डा. आशा सिंह)

प्रा. प्रहरी अतिरिक्त महानिरीक्षक