




नेपाल सरकार
गृह मन्त्रालय

नेपाल प्रहरी अस्पताल



फोन: { ०१-४४१२४३०
०१-४४१२५३०

Ext: { ०१-४४१२५३०
१२९

फो. व. नं. २११४७

फ्याक्स: ०१-४३५१५३

महाराजगंज, काठमाण्डौ ।

मिति: २०७७।०६।११ गते ।

पत्र संख्या: २०७७।०७८

प्रा.प्र.सं. ED

च.नं. १०१७

विषय : बोलपत्र सम्बन्धी कागजातहरूको संशोधन सम्बन्धमा ।

विश्व व्यापी रूपमा फैलिएको महामारी Coronavirus (COVID -19) को प्रकोप दैनिक रूपमा बढ्दै गईरहेकोले प्रत्यक्ष रूपमा खटिने नेपाल प्रहरीका सुरक्षाकर्मीहरू लगायत अन्य बिरामीहरू समेतको स्वास्थ्य हितलाई ध्यानमा राखि तत्काल सार्वजनिक सुरक्षा, हित तथा सामुदायिक स्वास्थ्यमा संकट पर्न गई थप धनजनको क्षतिलाई ब्यबस्थापन एवं न्युनिकरण गर्न यस अस्पतालमा भेण्टिलेटर सहितको २० वटा आईसियू बेडको Isolation with ICU Care तत्काल तयार गर्नु पर्ने भएकोले सार्वजनिक खरिद ऐन २०६३ (संसोधन सहित) को दफा ६६ र सार्वजनिक खरिद नियमावली २०६४ (संसोधन सहित) को नियम १४५ मा ब्यवस्था भए बमोजिम उक्त वार्डको स्थापनाको लागि अति आवश्यक बिभिन्न मेडिकल मेशिनरी तथा औजारहरू लगायतका सामग्रीहरू विशेष परिस्थितिमा आकस्मिक रूपमा तत्काल खरिद गर्नको लागि मिति २०७७।०६।०६ गतेको गोरखापत्र राष्ट्रिय दैनिकमा प्रकाशित सूचना बमोजिमको बोलपत्र सम्बन्धी कागजातहरूमा मिति २०७७।०६।११ गते १३:०० बजे यस अस्पतालमा आयोजना गरिएको बोलपत्र पेश गर्नु पूर्वको बैठक (Pre-Bid Meeting) मा भएको छलफल तथा बोलपत्रदाताहरूको तर्फबाट प्राप्त भएको सुझाव समेतलाई ध्यान दिई देहाय बमोजिम संशोधन/परिमार्जन गरिएकोले सम्बन्धित सबैको जानकारीको लागि अनुरोध गरिएको छ ।

साथै बोलपत्र दाखिला गर्ने तथा बोलपत्र खोल्ने मिति र समय यथावत रहेको व्यहोरा समेत जानकारीको लागि अनुरोध गरिएको छ ।

निम्न:

संशोधन गरिएको मेशिनरी उपकरणको सि.नं.	मेशिनरी उपकरणको नाम	बिबरण		
		बुँदा नं.	अस्पतालले निर्धारण गरेको स्पेशिफिकेशन	
3	ICU Ventilator for Adult, Pediatric and Neonatal.	4.3	The air passed through the compressor or turbine shall be cleaned by a HEPA filter before delivering to patient.	The air passed through the compressor shall be cleaned by a HEPA filter before delivering to patient.
		4.6	Microprocessor controlled gas delivery system with integrated air pump or a matching medical air compressor	Microprocessor controlled gas delivery system with integrated medical air compressor
		4.8	Battery back-up time: 4 to 6 Hours	Battery back-up time: 1 to 3 Hours

4	ICU Ventilator for Adult and Pediatric	4.3	The air passed through the compressor or turbine shall be cleaned by a HEPA filter before delivering to patient.	The air passed through the compressor or turbine shall be cleaned by a HEPA filter before delivering to patient. (Preferred Medical graded air Compressor)
		4.8	Battery back-up time: 4 to 6 Hours	Battery back-up time: 1 to 3 Hours
5	ICU Patient Monitor	4.1	High resolution colour flat panel non-reflective LED touch screen: 10" or more display size for minimum of 8 channel waveforms display.	High resolution colour flat panel non-reflective LED touch screen: 12" or more display and minimum of 10 channels wave from display.
		4.6	With powerful data storage for up to 48 hours of full disclosure, 1200 hours of tabular and graphic trends and 1600 NIBP measurements.	With powerful data storage for up to 48 hours of full disclosure, 1000 hours or more of tabular and graphic trends and 1000 NIBP or more measurements.
		4.21	Should have Lithium Ion battery to allow up to 4 hours (optional 8 hours) for continuous monitoring	Should have Lithium Ion battery to allow 2 to 4 hours (optional 8 hours) for continuous monitoring
7	Syringe pump	STRUCTURE:	Weight no more than 1.95 kg	Preferred touch screen display
		GENERAL FEATURES	Preset Volume(VTBI): 0.1-9999ml, minimum step 0.1ml	Preset Volume(VTBI): 0.01-999ml, minimum step 0.1ml
			Preset bolus volume: 0.1-9999ml	Preset bolus volume: 0.01-999ml
		SAFTY SPECIFICATION	Water-Proof Grade : IP34	Remove
8	Infusion Pump	STRUCTURE:	Weight no more than 1.6 kg	Preferred touch screen display
		GENERAL FEATURES	Preset Volume(VTBI): 0.1-9999ml	Preset Volume(VTBI): 0.01-999ml, minimum step 0.1ml
			Preset bolus volume: 0.1-9999ml	Preset bolus volume: 0.01-999ml
		SAFTY SPECIFICATION	Water-Proof Grade : IP34	Remove

कार्यार्थ एवं बोधार्थ:

श्री प्रि-विड मिटिडमा सहभागी सम्पूर्ण आपूर्तिकर्ताहरु ।
 श्री नेपाल प्रहरी अस्पताल, कम्प्यूटर शाखा: यस अस्पतालको
 Website मा राखिदिन हुन ।
 श्री नेपाल प्रहरी अस्पताल, सूचना पाटी ।

Nish
 मेडिकल डाईरेक्टर
 (डा. आशा सिंह)
 प्रा. प्रहरी अतिरिक्त महानिरीक्षक