



नेपाल सरकार
गृह मन्त्रालय
नेपाल प्रहरी अस्पताल
महाराजगंज, काठमाडौं

Handwritten signature

High end Video Endoscopy संग सम्बन्धीत मेशिनरी औजार आपूर्ति तथा जडान गर्ने सम्बन्धी सूचना

(प्रथम पटक प्रकाशित मिति २०७९।१०/२६ गते ।)

सूचना नं. १८/०७९/०८०

यस अस्पतालको मेडिसिन विभागमा रहेको Pentax Company कम्पनीको High end Video Endoscopy मेशिनको HD Pediatric Gastroscopy 1set थप गरि यस अस्पतालबाट सेवा लिने प्रहरी कर्मचारी, कर्मचारी परिवार, भु.पु. प्रहरी/परिवार तथा सर्व साधारण जनतालाई थप सेवा दिनको निमित्त खरिद तथा जडान गर्नको लागि Pentax Company कम्पनीको नेपाल स्थित आधिकारीक विक्रेताले मिति २०७९/११/०३ गते १७:०० वजे भित्र देहाय वमोजिमको कागजात सहित यस अस्पतालको प्रबन्ध शाखामा सम्पर्क राख्न हुन अनुरोध गरिएको छ ।

पेश गर्नुपर्ने कागजात :

१. उत्पादक कम्पनीले प्रदान गरेको आधिकारीक विक्रि अनुमति पत्र (Authorization Letter)
२. फर्म दर्ता तथा नविकरण गरेको प्रमाणपत्रको प्रमाणित प्रतिलिपी
३. मू.अ.कर दर्ता प्रमाण पत्रको प्रमाणित प्रतिलिपी
४. आ.व.०७८/०७९ को करचुक्ता प्रमाणपत्रको प्रमाणित प्रतिलिपि
५. खरिद गरिने मेशिनको प्राविधिक विवरण तथा दररेट
६. यस्तै प्रकृतिको माल सामान नेपालको कुनै अस्पतालमा आपूर्ति गरेको भए सो को विक्रि रेट
७. खरिद गरिने मेशिनरी औजारको माग Specification नेपाल प्रहरी अस्पतालको Web-site: <https://nph.nepalpolice.gov.np> मा राखिएको व्यहोरा जानकारीको लागि अनुरोध गरिएको छ ।

नेपाल प्रहरी अस्पताल
प्रबन्ध शाखा



dit
 Medical Director
 Dr. Asha Singh
 A.I.G.F

SN	Purchaser's Specifications		Bidders Compliance Sheet		
			Yes/No	Ref Docs Page No,	Remarks
	Manufacturer				
	Brand				
	Type/Model				
	Country of origin				
	Project Requirements:				
	The supplier must provide below listed products:				
	HD Paediatric Gastroscope 1 Set				
	Device Specification:				
1.	HD Video Gastroscope 1 Set				
	Technical Specifications				
1.1	It should have full screen crisp, clear image.				
1.2	It should support with image enhancement technology.				
1.3	The direction of view should be forward				
1.4	Field of view should be 140 Degree				
1.5	Tip angulation should be: Up 210 Deg Down 120 Deg Left 120 Deg Right 120 Deg				
1.6	Distal End should be 5-6 mm				
1.7	Instrument channel should be 2 mm or more				
1.8	Depth of field should be 3-100 mm with close focus technology to identify pit Patten.				
1.9	Working length 1050 mm or more				
1.10	Total length should be 1350 mm or more				
1.11	It should have reusable suction Valves and inlet seal				
1.12	The PVE connector should be rotatable to 180 degree to prevent damage of light guide fiber bundle.				
1.13	The scope should be compatible with existing processor in the Gastro Department.				
1.14	Backup of all the equipment should be provided in case of machine failure by the bidder. Commitment letter should be submitted for the same.				
2.	TERM AND CONDITION				
2.1	-User's and Maintenance training to be provided on site. -During warranty period supplier must ensure corrective/breakdown maintenance whenever required. -Inspections to verify the compliance of the offered goods as				



Nil
Medical Director
Dr. Asha Si
A.I.G.F

<p>per specifications will be conducted by the technical team appointed by the Purchaser.</p> <ul style="list-style-type: none">-Must submit the original brochure/catalogue/supporting documents with tender form.-User (Operating) and Technical/Maintenance manuals to be supplied in English.-Certification of calibration and inspection from factory-Certification of manufacturer Guaranteeing the availability of all spare parts for next 10(ten) year. <p>ISO Certified. CE and US FDA Certified.</p> <p>Warranty: Three years Complete warranty (Spare Parts and Service) and another two years service warranty (free maintenance service) from the installation date.</p>			
---	--	--	--