



पत्र संख्या: २०७९।०८०

प्रा.प्र.सं.

नेपाल सरकार
गृह मन्त्रालय

नेपाल प्रहरी अस्पताल



फोन: { ०१-४४१२४३०
०१-४४१२५३०

Ext.: { ०१-४४१२५३०
१२९

पो.ब.नं. २११४७

फ्याक्स: ०१-४३५१५३

महाराजगंज, काठमाण्डौ।

मिति: २०७९।१०।१२ गते।

विषय : विधुतिय बोलपत्र सम्बन्धी कागजातको संशोधन सम्बन्धमा।

उपरोक्त विषयमा यस अस्पतालको लागि आ.ब.२०७९/०८० मा आवश्यक पर्ने सर्जरी विभाग र डेण्टल विभागको लागि खरिद गरिने निम्नानुसारका मेशिनरी उपकरण तथा औजारहरु विधुतिय बोलपत्रको माध्यमबाट खरिद गर्न मिति २०७९।०९।१८ गते सोमवारको नेपाल समाचार पत्र राष्ट्रिय दैनिक मा प्रकाशित सुचना नं. १४/०७९/८० बमोजिमको बोलपत्र सम्बन्धी कागजातहरुमा मिति २०७९।१०।०६ गते शुक्रवार १२:०० बजे यस अस्पतालमा आयोजना गरिएको बोलपत्र पेश गर्नु पूर्वको बैठक (Pre-Bid Meeting) मा भएको छलफल तथा बोलपत्रदाताहरुको तर्फबाट भएको सुझाव एवं अस्पतालको आवश्यकता समेतलाई ध्यान दिई केहि बुदामा संशोधन/परिमार्जन गरिएकोले सम्बन्धित सबैको जानकारीको लागि अनुरोध गरिएको छ।

साथै बोलपत्र दाखिला गर्ने तथा बोलपत्र खोल्ने मिति र समय यथावत रहेको व्यहोरा समेत जानकारीको लागि अनुरोध गरिएको छ।

बोधार्थ एवं वितरण:

श्री प्रि-विड मिटिङमा सहभागी सम्पूर्ण आपूर्तिकर्ताहरु।

श्री नेपाल प्रहरी अस्पताल कम्प्यूटर शाखा, यस अस्पतालको

Website मा राखिदिन हुन।

श्री नेपाल प्रहरी अस्पताल, सूचना पाटी।


मेडिकल डाईरेक्टर

(डा. आशा सिंह)

प्रा. प्रहरी अतिरिक्त महानिरीक्षक



Government of Nepal
Ministry of Home Affairs
Nepal Police Hospital
Nepal Police Hospital, Maharajgunj, Kathmandu
Invitation for Electronic Bids
Notice No: NPH/NCB/14/079/080

hil
Medical Director
Dr. Asha Singh
A.I.G.P

Date of Publication: 2nd January 2023 (18th Push 2079 BS)

1. Nepal Police Hospital invites **Bids** from the eligible bidders for the procurement of following goods under **National Competitive bidding procedures**. The name and identification of the contact are as follows.

SN	Descriptions	Contract No.	Qty.	Bid Security Amount (NRs)	Bid Document Fee (NRs)
1.	4K Laparoscopy Tower System	NPH/NCB/14/078/079/01	1 Sets	3,12,600.00	5,000.00
2.	C-arm Machine	NPH/NCB/14/078/079/02	1 Sets	2,97,000.00	
3.	Dental Operation Microscope	NPH/NCB/14/078/079/03	1 Sets	1,10,000.00	

2. Eligible Bidders may obtain further information from Nepal Police Hospital, Account/Procurement Section, Maharajgunj, Kathmandu, tel: 01-4412430, 01-4412530, Ext. No. 125/128 and email address: nph.procurementsec@nepalpolice.gov.np.
3. Bidding documents is available online and can be downloaded from e-GP system: www.bolpatra.gov.np/egp.
Pre-bid meeting shall be held at Nepal Police Hospital Maharajgunj, Kathmandu on **20th January 2023 (6th Magh 2079 BS) at 12:00 hours**.
4. If the any amendment of bids documents, notice shall be posted on the notice board and website of the Nepal Police Hospital, Maharajgunj.



hil

Medical Director
Dr. Asha Singh
A.I.G.P

सि.नं	मेशिनरी उपकरणको नाम	बुँदा नं.	अस्पतालले निर्धारण गरेको स्पेशिफिकेशन	संशोधन तथा थप भएको विवरण
1.	4 K Laparoscopy Tower set with instruments	2A	The laparoscopic surgery camera, video processor, Light source, monitor, Recorder,insufflator along with trolley must have individual/Composite components as given below. Complete system must be from same manufacture.	The laparoscopic surgery camera, video processor, Light source, monitor, Recorder, insufflator along with trolley must have individual/Composite components as given below. Camera system should
		4A	4K Camera Head CCD/ CMOS 3-Chip/CMOS full 4K Resolution with different feature categories. Image Resolution full 4K with greater or equal to 4096 x 2160 pixels.	4K Camera Head CMOS 3-Chip full 4K Resolution with different feature categories. Image Resolution full 4K with greater or equal to 3840 x 2160 pixels.
			Image sensor : 3x1/3" CCD/CMOS	Image sensor: 3 chip CMOS
			The unit shall come with CCU, camera head with 3m or more cable, mains cable, Y/C cable, BNC cable, HDMI - DVI cable, HDMI cable	The unit shall come with CMOS, camera head with 3m or more cable, mains cable, Y/C cable, BNC cable, HDMI - DVI cable, HDMI cable or 1X3G-SDI, 2XDVI-D
			Modern and clearly arranged On-Screen menu.	The point has been removed.
			Should have user friendly 6"touch screen system or feather touch button with On-Screen menu.	The point has been removed.
			Should have at least 2 or more camera head buttons can operate 4 programs or more (user selectable).	Should have at least 3 or more camera head buttons and operate 6 programs or more (user selectable).
			Camera head should have Autoclavable, immersible and gas sterilizable	camera head should be immersable and gas sterilizable
			4B	300W xenon / LED light source
		4E	Must have an adjustable flow rate up to 45l/min and a pressure range adjustable	Must have an adjustable flow rate up to 40L/min and a pressure range adjustable between 0-25



dit

Medical Director
Dr. Asha Singh

			between 0-25 mmHg	mmHg
		7A	Must submit USFDA and European CE (93/42 EEC Directives). Certified.	Must submit USFDA and European CE (93/42 EEC Directives). Certified for camera system and instruments.
2.	C-arm Machine	A2	Output power should be 2.5 KW or more.	Output power should be at least 2 KW or more.
		A4	Fluoroscopic mA lower limit : 0.1 mA upper limit : 10 mA or more	Fluoroscopic mA lower limit : 0.2 mA upper limit : 16 mA or more
		A5	Tube head having dual focus stationary anode X-Ray tube of focal spot 0.6mm X 1.44mm (small focus spot) & large focus spot (1.4 mm x 1.4mm) should be provided.	Tube head having single or dual focus stationary anode X-Ray tube of focal spot 0.6mm and or 0.6x1mm.
		B1	Rotation: ± 180 Degrees or more.	Rotation: ± 135 Degrees or more.
3.	Dental Operating Microscope	4.24	Should be supplied with 32-inch Full HD Display Monitor on stand for displaying images/video during recording.	Should be supplied with 32-inch medical grade Full HD Display Monitor on stand for displaying images/video during recording.