



पत्र संख्या: २०७९।०८०  
प्रा.प्र.सं.

नेपाल सरकार  
गृह मन्त्रालय  
नेपाल प्रहरी अस्पताल



च.नं. ४०८६

फोन: { ०१-४४१२४३०  
०१-४४१२५३०

Ext: { ०१-४४१२५३०  
१२९

पो.ब.नं. २११४७

फ्याक्स: ०१-४३५१५३

महाराजगंज, काठमाण्डौ।

मिति: २०७९।१२।०९ गते।

**विषय : विद्युतिय बोलपत्र सम्बन्धी कागजातको संशोधन सम्बन्धमा।**

उपरोक्त विषयमा यस अस्पतालको लागि आ.व. २०७९/०८० मा खरिद गर्न लागिएको MRI 3.0 Tesla Machine खरिद गर्नु पर्ने भएकोले विद्युतिय बोलपत्रको माध्यमबाट खरिद गर्न मिति २०७९/१०/२९ गते आईतवार अन्नपूर्ण पोष्ट राष्ट्रिय दैनिक पत्रिकामा प्रकाशित सूचना नं. NPH/ICB/19/079/080 बमोजिमको बोलपत्र सम्बन्धी कागजातहरूमा मिति २०७९।११।२३ गते १२:०० बजे यस अस्पतालमा आयोजना गरिएको बोलपत्र पेश गर्नु पूर्वको बैठक (Pre-Bid Meeting) मा भएको छलफल तथा बोलपत्रदाताहरूको तर्फबाट भएको सुझाव एवं अस्पतालको आवश्यकता समेतलाई ध्यान दिई केहि बुदामा संशोधन/परिमार्जन गरिएकोले सम्बन्धित सबैको जानकारीको लागि अनुरोध गरिएको छ।

साथै बोलपत्र दाखिला गर्ने तथा बोलपत्र खोल्ने मिति र समय निम्न अनुसार रहेको व्यहोरा समेत जानकारीको लागि अनुरोध गरिएको छ।

**निम्न:**

बोलपत्र दाखिला गर्ने मिति र समय : २०७९।१२।२१ (4th April 2023) दिनको १२:०० बजे

बोलपत्र खोल्ने मिति र समय : २०७९।१२।२१ (4th April 2023) दिनको १३:०० बजे


**बोधार्थ एवं वितरण:**

श्री प्रि-विड मिटिङमा सहभागी सम्पूर्ण आपूर्तिकर्ताहरू।

श्री नेपाल प्रहरी अस्पताल कम्प्युटर शाखा, यस अस्पतालको

Website मा राखिदिन हुन।

श्री नेपाल प्रहरी अस्पताल, सूचना पाटी।

  
मेडिकल डाईरेक्टर

( डा. आशा सिंह )

प्रा. प्रहरी अतिरिक्त महानिरीक्षक



सि.नं	मेशिनरी उपकरणको नाम	बुँदा नं.	बिबरण	संशोधन तथा थप भएको विवरण
			अस्पतालले निर्धारण गरेको स्पेशिफिकेशन	
1.	3.0 Tesla MRI Machine	1.	A new high end 3T MRI with active shielded superconductive magnet.	A new high end 3T MRI with active shielded superconductive magnet. <b>No refurbished/ ETN-Equivalent-to – Latest/New Parts should be used in the MRI system offered. The date of manufacture should be clearly mentioned directly from the manufacturer with authorized signature and stamped.</b>
		2.B	Auto shim (global and oxel shim) should take minimum time to shim the magnet with patient in position. High order (2 <sup>nd</sup> order) Shim should be offered as standard.	<b>High order or Auto-Shim should be offered as standard.</b>
		7.C.	Total hard disk memory to be sufficient to store at least 20,00,000 images of 256x256 matrix data size.	Total hard disk memory to be sufficient to store at least <b>12,00,000</b> images of 256x256 matrix data size.
		17.H	Nos. UPS : Compatible UPS for entire MRI system including console and workstation with at least 30 minutes back up Necessary Air Conditioning (as number mentioned in BOQ) with Dehumidification should be supplied according to the manufacturer's specification for control console, workstation, liquid helium, chiller, computer system, printer and Batteries for UPS and accessories.	Nos. UPS : Compatible UPS for entire MRI system including console and workstation with at least 30 minutes back up Necessary Air Conditioning with Dehumidification should be supplied <b>as required by hospital</b> for control console, workstations, <b>technical and electronic cabinet room and other room as required by the hospital should be provided.</b> Also, liquid helium, chiller, computer system, printer and Batteries for UPS and accessories <b>should be provided as required.</b> <b>The hospital won't be responsible in any matter for any issues.</b>